

ielab

Proficiency Testing Schemes

Atmospheric pollution – 2012

Additional information

Alicante, December 26th 2012

Dear participants,

According to the values reported for the 2012 Atmospheric pollution ielab Scheme, the results obtained for participant codes **0103, 0261, 0430, 0601, 4165, 6016** and **6963** in each of the three rounds are shown as follows.

RESULTS PRESENTATION

- For each parameter the participants' results are shown in a table, including mean and standard deviation for each laboratory. The **consensus** value X with the standard deviation for proficiency assessment (SDPA σ_p) and the uncertainty are also included in this table. In order to evaluate the performance of the participants, Z-score value appears in the last column.
- The laboratories that have not reported results according the round instructions (less than three replicates, or non-numerical values) have not been included in the statistical treatment and appear shaded.
- The eliminated results after applying the statistical treatment are shown shaded. The laboratories, whose intra-laboratory deviation is higher than 0.5 times the standard deviation for proficiency assessment, appear with their deviation cell boxed.

METHODS COMPARISON TABLE

For each parameter and in addition to the general calculations, a study about interactions between methods and results has been developed. The methods comparison table is also annexed, where it is shown for each parameter:

- Method of reference.
- The number of laboratories with acceptable values that has been used that method.
- The robust mean and the robust standard deviation for those results.

The group of laboratories that declared the method "others methods" has not taken into consideration in this comparison table. Nor laboratories that declared methods not included for each parameter in the method sheet.

It has only been considered methods used for more than 5 laboratories.

Full information of the scheme can be found in the corresponding Round main report.

FILTERS – Round I

ARSENIC ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.18	D1	R2	12.2	11.4	11.9	11.83	0.40	11.43 12.24	0.23
0261	M32	0	D1	R1	11.28	11.18	8.9	10.45	1.35	9.11 11.80	-0.19
0430	M52	<0.25	D1	R1	9.29	9.65	10.2	9.71	0.46	9.26 10.17	-0.41
0601					13.2	13	13.1	13.10	0.10	13.00 13.20	0.61
4165											
6016											
6963	M52	0	D1	R1	9.8	10.1	10	9.97	0.15	9.81 10.12	-0.34

N	61	Consensus value	11.08	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ	3.32	M32	5	10.45	0.95
		Robust devest.	1.25	M51	32	11.22	1.02
		Uncertainty	0.160	M52	23	11.00	1.37

COBALT ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.005	D1	R2	8.76	8.14	8.46	8.45	0.31	8.14 8.76	0.06
0261	M32	0	D1	R1	7.8	7.97	6	7.26	1.09	6.17 8.35	-0.42
0430	M52	0.32	D1	R1	7.46	7.7	7.97	7.71	0.26	7.45 7.97	-0.24
0601					9.4	9.2	9.3	9.30	0.10	9.20 9.40	0.40
4165	M25	0			7.5	7.9	7.5	7.63	0.23	7.40 7.86	-0.27
6016		<0.17			8.900	8.100	8.300	8.43	0.42	8.02 8.85	0.06
6963	M52	0	D1	R1	8.5	8.7	8.8	8.67	0.15	8.51 8.82	0.15

N	59	Consensus value	8.30	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ	2.49	M51	31	8.37	0.73
		Robust devest.	0.77	M52	24	8.22	0.68
		Uncertainty	0.100				

MANGANESE ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	0.024	D1	R2	21.9	20.4	21.1	21.13	0.75	20.38 21.88	0.00
0261											
0430	M52	<0.004	D1	R1	19.9	20.6	20.97	20.49	0.54	19.95 21.03	-0.10
0601					24.2	23.3	23.9	23.80	0.46	23.34 24.26	0.42
4165	M25	0			20	20.4	20	20.13	0.23	19.90 20.36	-0.16
6016		0.143			21.929	20.464	21.036	21.14	0.74	20.40 21.88	0.00
6963	M52	0	D1	R1	21.1	21.6	21.9	21.53	0.40	21.13 21.94	0.06

N	59	Consensus value	21.13	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ	6.34	M51	30	21.24	1.63
		Robust devest.	1.26	M52	26	21.08	0.91
		Uncertainty	0.164				

NICKEL ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	0.074	D1	R2	20.2	19	19.5	19.57	0.60	18.96 20.17	0.10
0261	M32	0.24	D1	R1	17.14	17.5	14.25	16.30	1.78	14.52 18.08	-0.47
0430	M52	<0.039	D1	R1	17.08	17.53	18.19	17.60	0.56	17.04 18.16	-0.24
0601					24.3	22.6	20.8	22.57	1.75	20.82 24.32	0.63
4165	M25	0			17.3	17.9	17.1	17.43	0.42	17.02 17.85	-0.27
6016		6.385			20.692	19.231	20.231	20.05	0.75	19.30 20.80	0.19
6963	M52	0	D1	R1	19.5	19.3	19.5	19.43	0.12	19.32 19.55	0.08

N	64
---	----

Consensus value	18.99	METHOD	LAB No	ROB MEAN	ROB SD
SDPA op	5.70	M51	33	19.15	1.41
Robust devest.	1.52	M52	25	18.86	1.29
Uncertainty	0.190				


VANADIUM ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.018	D1	R2	9.89	9.3	9.56	9.58	0.30	9.29 9.88	0.02
0261											
0430	M52	<0.034	D1	R1	9.84	10.18	10.31	10.11	0.24	9.87 10.35	0.21
0601					8.3	7.7	8.5	8.17	0.42	7.75 8.58	-0.47
4165											
6016											
6963	M52	0	D1	R1	8.7	8.8	9.1	8.87	0.21	8.66 9.07	-0.23

N	57
---	----

Consensus value	9.51	METHOD	LAB No	ROB MEAN	ROB SD
SDPA op	2.85	M51	30	9.45	0.84
Robust devest.	0.82	M52	26	9.65	0.73
Uncertainty	0.109				

NOTE- In the first round, laboratory code 6016 results have not been considered in the general statistics because they were received after the results submission deadline.



FILTERS – Round II

ANTIMONY ($\mu\text{g/filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	0.46	D1	R2	27.8	29.2	28.7	28.57	0.709	27.86 29.28	-0.23
0261											
0430	M52	0.38	D1	R3	20.11	19.13	19.34	19.53	0.516	19.01 20.04	-1.22
0601	M52	<1		R6	27.3	26.3	26.1	26.57	0.643	25.92 27.21	-0.45
4165											
6016											
6963	M52	0	D1	R3	23.5	25	24	24.17	0.764	23.40 24.93	-0.71

N	56
---	----

Consensus value	30.73	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ	9.219	M51	27	32.40	3.386
Robust devest.	4.258	M52	25	29.61	3.902
Uncertainty	0.569				

CADMIUM ($\mu\text{g/filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.02	D1	R2	14.6	14.6	14.2	14.47	0.231	14.24 14.70	0.11
0261	M32	0.06	D1	R1	13.04	12.39	12.53	12.65	0.342	12.31 13.00	-0.32
0430	M52	<0.038	D1	R3	13.03	12.6	12.65	12.76	0.235	12.52 13.00	-0.30
0601	M52	<1		R6	12.5	12.2	12.2	12.30	0.173	12.13 12.47	-0.41
4165											
6016	M25	<0.05	D1	R1	12.78	12.97	11.55	12.43	0.771	11.66 13.20	-0.37
6963	M52	0	D1	R3	12.5	12.9	12.5	12.63	0.231	12.40 12.86	-0.33

N	68
---	----

Consensus value	14.00	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ	4.200	M51	32	14.22	1.249
Robust devest.	1.301	M52	29	13.84	1.258
Uncertainty	0.158				

CHROMIUM ($\mu\text{g/filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	0.24	D1	R2	22	22	21.4	21.80	0.346	21.45 22.15	0.09
0261	M32	0.34	D1	R1	21.3	21.6	21.9	21.60	0.300	21.30 21.90	0.06
0430	M52	<0.042	D1	R3	21.65	19.68	19.92	20.42	1.075	19.34 21.49	-0.13
0601	M52	<1		R6	18.5	18.5	18.5	18.50	0.000	18.50 18.50	-0.43
4165											
6016	M25	6.36	D1	R1	19.51	20.25	19.93	19.90	0.371	19.53 20.27	-0.21
6963	M52	0	D1	R3	19.5	20.2	19.7	19.80	0.361	19.44 20.16	-0.22

N	64
---	----

Consensus value	21.22	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ	6.366	M51	29	21.73	2.371
Robust devest.	1.944	M52	29	20.87	1.708
Uncertainty	0.243				

TIN ($\mu\text{g}/\text{filter}$)


LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.02	D1	R2	17	16.7	16.3	16.67	0.351	16.32 17.02	-0.07
0261											
0430	M52	<0.028	D1	R3	14.31	13.43	13.09	13.61	0.630	12.98 14.24	-0.67
0601	M52	<1		R6	90	21.7	18.5	43.40	40.388	3.01 83.79	5.15
4165											
6016											
6963	M52	0	D1	R3	13.8	14.5	14.1	14.13	0.351	13.78 14.48	-0.57

N	49	Consensus value	17.05	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	5.115	M51	25	17.66	1.688
		Robust devest.	1.935	M52	24	16.40	1.997
		Uncertainty	0.276				

MERCURY ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103				R2							
0261											
0430	M52	0.03	D1	R3	0.73	0.67	0.69	0.70	0.031	0.67 0.73	-3.12
0601	M45			R2	10.4	12.9	12.4	11.90	1.323	10.58 13.22	0.32
4165											
6016											
6963	M45	no blank	D1	R1	10571	10266	10458	10431.67	154.196	10277.47 10585.86	3198.53

N	43	Consensus value	10.86	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	3.258	M45	7	11.88	2.362
		Robust devest.	1.733	M48	5	11.58	1.008
		Uncertainty	0.264	M51	11	10.77	1.698
				M76	12	10.32	1.453



FILTERS – Round III
COPPER ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.03	D1	R2	27.4	28.5	29	28.30	0.819	27.48 29.12	-0.20
0261											
0430	M52	0	D1	R1	28.23	28.16	28.28	28.22	0.060	28.16 28.28	-0.21
0601	M52	<2	D1	R2	32.1	29.8	26.7	29.53	2.710	26.82 32.24	-0.07
4165											
6016	M25	<1	D1	R1	28.3	28	26.32	27.54	1.067	26.47 28.61	-0.29
6963	M52	0.1	D1	R1	24.6	26.7	23.7	25.00	1.539	23.46 26.54	-0.57

N	55	Consensus value	30.14	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ	9.042	M25	3	29.25	1.726
		Robust devest.	2.103	M51	26	30.20	2.357
		Uncertainty	0.284	M52	24	30.12	1.767

LEAD ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	0.06	D1	R2	23.9	23.7	24.7	24.10	0.529	23.57 24.63	0.03
0261	M25	0	D1	R1	21.75	21.75	20.05	21.18	0.981	20.20 22.16	-0.37
0430	M52	0	D1	R1	21.6	21.47	21.79	21.62	0.161	21.46 21.78	-0.31
0601	M52	<2	D1	R2	20	23.2	23	22.07	1.793	20.27 23.86	-0.25
4165											
6016	M25	<3.2	D1	R1	19.4	19.15	17.18	18.58	1.216	17.36 19.79	-0.74
6963	M52	0	D1	R1	18.4	19.7	22.8	20.30	2.261	18.04 22.56	-0.50

N	61	Consensus value	23.86	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ	7.158	M51	31	24.12	1.730
		Robust devest.	2.191	M52	23	23.85	2.126
		Uncertainty	0.280				

SELENIUM ($\mu\text{g}/\text{filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	0.3	D1	R2	31.9	33.4	34.2	33.17	1.168	32.00 34.33	0.25
0261											
0430	M52	0	D1	R1	24.04	24.4	24.12	24.19	0.189	24.00 24.38	-0.72
0601	M52	<2	D1	R2	27	32.7	30	29.90	2.851	27.05 32.75	-0.10
4165											
6016											
6963	M52	0.1	D1	R1	24.6	25.9	26.4	25.63	0.929	24.70 26.56	-0.56

N	50	Consensus value	30.84	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ	9.252	M51	25	31.72	2.985
		Robust devest.	4.597	M52	21	30.13	5.362
		Uncertainty	0.650				

THALLIUM ($\mu\text{g/filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.005	D1	R2	9.8	11.5	11.8	11.03	1.079	9.95 12.11	-1.38
0261											
0430	M52	0.15	D1	R1	16.52	16.32	16.25	16.36	0.140	16.22 16.50	-0.44
0601	M52	<2	D1	R2	15.6	18.2	19	17.60	1.778	15.82 19.38	-0.22
4165											
6016											
6963	M52	0.1	D1	R1	15.2	15.6	17.7	16.17	1.343	14.82 17.51	-0.47

N	54
---	----


Consensus value	18.84	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ_p	5.652	M51	29	18.97	1.497
Robust devest.	1.536	M52	22	18.81	1.302
Uncertainty	0.209				

ZINC ($\mu\text{g/filter}$)

LAB	MET	BLANK	DIG	TYPE	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.2	D1	R2	18.9	19.2	20	19.37	0.569	18.80 19.94	-0.03
0261											
0430	M52	1.58	D1	R1	21	20.3	20.43	20.58	0.372	20.20 20.95	0.18
0601	M52	<2	D1	R2	15.5	18.7	18	17.40	1.682	15.72 19.08	-0.36
4165											
6016	M25	2.37	D1	R1	21.96	22.07	20.26	21.43	1.015	20.42 22.44	0.33
6963	M52	0.3	D1	R1	15.4	16.5	16.9	16.27	0.777	15.49 17.04	-0.56

N	52
---	----

Consensus value	19.52	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ_p	5.856	M51	22	19.41	2.687
Robust devest.	2.046	M52	24	19.49	1.427
Uncertainty	0.284				



ABSORPTION SOLUTIONS – Round I

ANTIMONY ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<20	224	230	233	229	4.58	224 234	-0.32
0261									
0430	M52	12.4	200.7	199.4	200.1	200	0.65	199 201	-0.92
0601			83.9						
4165									
6016									
6963	M52	0	243	224	212	226	15.63	211 242	-0.37

N	52	Consensus value	244	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	48.31	M51	28	250	19.75
		Robust devest.	17.47	M52	24	239	17.13
		Uncertainty	2.4				

ARSENIC ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<2	94.1	102	94.2	97	4.53	92 101	0.13
0261	M32	0	128.8	123.5	118.4	124	5.20	118 129	1.43
0430	M52	<9.9	86.8	78.8	82.8	83	4.00	79 87	-0.54
0601			30.6						
4165									
6016									
6963	M52	0	75	79	75	76	2.31	74 79	-0.86

N	59	Consensus value	94	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	20.70	M32	5	89	25.37
		Robust devest.	8.22	M51	32	96	6.85
		Uncertainty	1.1	M52	22	93	7.88

CADMIUM ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.5	129	138	139	135	5.51	130 141	-0.18
0261	M32	0	245.2	245.6	201	231	25.64	205 256	2.97
0430	M52	<0.5	127.3	127.9	127.6	128	0.30	127 128	-0.43
0601			73.3						
4165	M25	0	140	143	140	141	1.73	139 143	0.01
6016		10.832	159.847	150.232	150.588	154	5.45	148 159	0.42
6963	M52	0	144	146	145	145	1.00	144 146	0.14

N	59	Consensus value	141	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	30.24	M51	31	141	6.41
		Robust devest.	7.91	M52	24	140	9.24
		Uncertainty	1.0				

COPPER ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<1	438	483	489	470	27.87	442 498	-0.24
0261									
0430	M52	<0.4	454.5	456.2	455.4	455	0.85	455 456	-0.41
0601			443.6						
4165									
6016									
6963	M52	0	507	513	501	507	6.00	501 513	0.18

N	58
---	-----------

Consensus value	491	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ_p	87.41	M51	30	498	22.07
Robust devest.	25.46	M52	25	485	27.09
Uncertainty	3.3				


HF (HYDROFLUORIC ACID) (mg/L)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103									
0261									
0430									
0601		0.00	0.86						
4165									
6016									
6963	M41	0	0.922	0.926	0.949	0.93	0.015	0.92 0.95	-0.26

N	50
---	-----------

Consensus value	0.97	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ_p	0.156	M36	22	0.96	0.10
Robust devest.	0.075	M41	26	0.98	0.06
Uncertainty	0.011				

NOTE- In the first round, laboratory code 6016 results have not been considered in the general statistics because they were received after the results submission deadline.



ABSORPTION SOLUTIONS – Round II

CHROMIUM ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.2	285.2	285.8	288	286	1.5	285 288	-0.30
0261	M32	17.7	290.3	288.3	282.3	287	4.2	283 291	-0.29
0430	M52	<2.8	336.3	334.3	343.1	338	4.6	333 343	0.59
0601	M52	<0.01	0.32	0.31	0.33	0.32	0.01	0.31 0.33	-5.22
4165									
6016	M25	<0.3	253.33	263.53	240	252	11.8	240 264	-0.88
6963	M52	0	333	344	333	337	6.4	330 343	0.57

N	65	Consensus value	304	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	58.1	M51	29	308	14.9
		Robust devest.	16.8	M52	30	304	14.1
		Uncertainty	2.1				

MANGANESE ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.0	133.9	133.7	136.3	135	1.4	133 136	-0.28
0261									
0430	M52	<0.2	136.4	134.1	144.1	138	5.2	133 143	-0.16
0601	M52	<0.01	14	14	14	14	0.0	14 14	-4.21
4165									
6016	M25	<0.1	128.18	117.07	131.61	126	7.6	118 133	-0.57
6963	M52	0	157	163	158	159	3.2	156 163	0.53

N	62	Consensus value	143	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	30.7	M51	29	145	7.1
		Robust devest.	7.2	M52	30	142	5.8
		Uncertainty	0.9				

LEAD ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<1.3	183	178.7	182.2	181	2.3	179 184	-0.11
0261	M32	0	242	235	236	238	3.8	234 241	1.36
0430	M52	<8.8	204.3	207.6	202	205	2.8	202 207	0.49
0601	M52	<0.05	0.6	0.5	0.4	0.5	0.1	0.4 0.6	-4.84
4165									
6016									
6963	M52	0	206	203	189	199	9.1	190 208	0.36

N	63	Consensus value	186	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	38.3	M51	29	188	9.3
		Robust devest.	11.5	M52	29	183	11.8
		Uncertainty	1.4				

VANADIUM ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<0.2	77.7	77.1	79.1	78	1.0	77 79	-0.19
0261									
0430	M52	<2.2	80.8	80.1	82.8	81	1.4	80 83	0.00
0601	M52	<0.05	0.07	0.08	0.07	0.07	0.01	0.07 0.08	-4.54
4165									
6016									
6963	M52	0	88	91	87	89	2.1	87 91	0.41

N	58
---	-----------


Consensus value	81	METHOD	LAB No	ROB MEAN	ROB SD
SDPA σ_p	17.9	M51	28	82	3.9
Robust devest.	4.5	M52	30	80	4.6
Uncertainty	0.6				

HCl (HYDROCHLORIC ACID) (mg/L)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103									
0261									
0430									
0601	M36	<1	100	100	100	100.0	0.00	100.0 100.0	-0.35
4165									
6016									
6963	M36	0	103	104	104	103.7	0.58	103.1 104.2	0.10

N	55
---	-----------

Consensus value	102.9
SDPA σ_p	8.19
Robust devest.	4.09
Uncertainty	0.55



ABSORPTION SOLUTIONS – Round III

COBALT ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	37.5	295.2	294.2	298.6	296	2.3	294 298	0.16
0261	M25	35.0	296	288	281	288	7.5	281 296	0.02
0430	M52	25.4	337.9	344.1	340	341	3.2	338 344	0.97
0601	M52	34000	283000						
4165									
6016	M25	50.88		294.5					
6963	M52	23.5	292	280	277	283	7.9	275 291	-0.08

N	58	Consensus value	287	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	55.4	M51	26	284	33.0
		Robust devest.	30.5	M52	26	287	25.7
		Uncertainty	4.0				

TIN ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	<2.8	104	105.5	104.9	105	0.8	104 106	0.17
0261									
0430	M52	0	118.6	108.1	107.3	111	6.3	105 118	0.47
0601	M52	<1000	96000						
4165									
6016		<105		300.71					
6963	M52	0	113	104	118	112	7.1	105 119	0.48

N	46	Consensus value	101	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	22.2	M51	23	101	5.1
		Robust devest.	5.5	M52	21	102	6.8
		Uncertainty	0.8				

NICKEL ($\mu\text{g/L}$)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	17.8	325.5	326.8	331.2	328	3.0	325 331	0.14
0261	M25	12.0	354	354	356	355	1.2	354 356	0.59
0430	M52	10.4	361.4	366.3	361.5	363	2.8	360 366	0.73
0601	M52	15000	310000						
4165									
6016	M25	<22		132.76					
6963	M52	15	337	320	320	326	9.8	316 335	0.11

N	58	Consensus value	319	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	60.6	M51	26	320	24.5
		Robust devest.	24.2	M52	25	316	17.6
		Uncertainty	3.2				

ZINC ($\mu\text{g/L}$)


LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103	M52	3.6	436.1	439	439.9	438	2.0	436 440	0.06
0261									
0430	M52	55.4	510.4	547.6	500.1	519	25.0	494 544	1.09
0601	M52	<1000	423000						
4165									
6016	M25								
6963	M52	2	265	250	256	257	7.5	249 265	-2.24

N	56	Consensus value	434	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	78.7	M51	26	430	45.4
		Robust devest.	33.9	M52	26	436	24.2
		Uncertainty	4.5				

SO₂ (SULPHUR DIOXIDE) (mg/L)

LAB	MET	BLANK	X1	X2	X3	MEAN	SD	INTERVAL	Z-SCORE
0103									
0261									
0430									
0601	M36	<0.01	500						
4165									
6016									
6963	M114	<0.1	436	423	432	430	6.7	424 437	-0.52

N	43	Consensus value	445	METHOD	LAB No	ROB MEAN	ROB SD
		SDPA σ_p	28.4	M36	31	450	24.7
		Robust devest.	26.5	M114	12	433	27.1
		Uncertainty	4.0				



To which the present document of additional information about the **ielab** Scheme Atmospheric Pollution 2012 is issued, for it to have its due effects, this day.

Estíbaliz Sastre González
 Technical Area Coordinator
e.sastre@ielab.es
www.ielab.es

